Roche Cobas C311

Fardis Laboratory

## MONTHLY CLINICAL CHEMISTRY

CYCLE 19 SAMPLE 10

## Explanation of codes used in this report

R - Results removed due to reconstitution error

N - No result returned

C - Result corrected

Authorised by: Stephen Doherty, RIQAS Manager

Issue No: I

Issue Date: 04/11/2022

Randox Laboratories Limited 55 Diamond Road CRUMLIN BT29 4QY Tel: +44 (0)28 9445 4399

Tel: +44 (0)28 9445 4399 Fax: +44 (0)28 9445 4398 Email: mail@riqas.com





Analyte	Mean for Comparison	Your Result	SDI	RMSDI	%DEV	RM%DEV	TS	RMTS	Performanc
Angiotensin Converting En	zyme (AC <b>E</b> ).009	47.000	1.44	Too Few	51.6	Too Few	N/A	N/A	
Acid Phos., Prostatic	43.580	No Result		Too Few		Too Few	N/A	N/A	
Acid Phos., Total	53.168	51.700	-0.23	Too Few	-2.8	Too Few	N/A	N/A	
Albumin	3.187	3.130	-0.33	Too Few	-1.8	Too Few	120	Γοο Few	
Alkaline Phosphatase	323.917	340.000	0.44	Too Few	5.0	Too Few	107	Γοο Few	
ALT (GPT)	128.360	130.000	0.14	Too Few	1.3	Too Few	120	Γοο Few	
Amylase, Total	293.921	297.000	0.10	Too Few	1.0	Too Few	120	Γοο Few	
AST (GOT)	132.159	133.800	0.14	Too Few	1.2	Too Few	120	Γοο Few	
Bilirubin, Direct	1.823	1.500	-1.13	Too Few	-17.7	Too Few	66	Γοο Few	
Bilirubin, Total	4.784	4.900	0.25	Too Few	2.4	Too Few	120	Γοο Few	
Calcium	13.073	12.950	-0.19	Too Few	-0.9	Too Few	120	Γοο Few	
Cholinesterase	5285.857	5267.000	-0.04	Too Few	-0.4	Too Few	120	Γοο Few	
Chloride	110.174	112.000	0.59	Too Few	1.7	Too Few	94 7	Γοο Few	
Cholesterol	299.553	297.000	-0.16	Too Few	-0.9	Too Few	120	Γοο Few	
CK, Total	507.021	456.000	-1.38	Too Few	-10.1	Too Few	58 7	Γοο Few	
Copper	154.194	173.000	0.57	Too Few	12.2	Too Few	96 7	Γοο Few	
Creatinine	4.210	4.150	-0.19	Too Few	-1.4	Too Few		Γοο Few	
GGT	165.192	163.000	-0.12	Too Few	-1.3	Too Few		Γοο Few	
Glucose	277.176	279.000	0.13	Too Few	0.7	Too Few		Γοο Few	
HDL-Cholesterol	119.034	100.000	-1.25	Too Few	-16.0	Too Few		Γοο Few	
Iron	213.408	210.000	-0.26	Too Few	-1.6	Too Few		Γοο Few	
Lactate	49.701	48.700	-0.36	Too Few	-2.0	Too Few		Γοο Few	
LD (LDH)	405.604	408.000	0.08	Too Few	0.6	Too Few		Γοο Few	
LDL-Cholesterol (Pilot)	157.411	155.000	-0.12	Too Few	-1.5	Too Few		Too Few	
Lipase	74.818	63.000	-1.01	Too Few	-1.5	Too Few		Γοο Few	
Lithium	1.920	1.830	-0.63	Too Few	-4.7	Too Few		Too Few	
Magnesium	4.232	4.200	-0.03	Too Few	-0.7	Too Few		Γοο Few	
Phosphate, Inorganic	7.204	7.200	-0.01	Too Few	-0. <i>1</i>	Too Few		Γοο Few	
Potassium	6.231	6.200	-0.15	Too Few	-0.1	Too Few		Γοο Few	
			-0.13	Too Few		Too Few			
Protein, Total	4.608	4.600			-0.2			Γοο Few	
Sodium	160.288	157.000	-0.94	Too Few	-2.1	Too Few		Γοο Few	
TIBC	228.097	No Result	0.00	Too Few	<b>^</b> 1	Too Few		Γοο Few	
Trig Total	251.361	251.000	-0.02	Too Few	-0.1	Too Few		Γοο Few	
UIBC	15.391	15.000	-0.09	Too Few	-2.5	Too Few		Γοο Few	
Urea	124.216	119.400	-0.51	Too Few	-3.9	Too Few		Γοο Few	
Uric Acid (Urate)	9.353	9.300	-0.09	Too Few	-0.6	Too Few		Γοο Few	
Zinc	236.889	258.000	0.51	Too Few	8.9	Too Few	101	Γοο Few	
				ISDI N/A		1%DEV N/A		S N/A	

END OF REPORT





Issue No: I Issue Date: 04/11/2022 Page 39